

## **Abstract**

A self expandable Nitinol stent was accidentally maldeployed when precise positioning was being done in a pre dilated 100% occluded left subclavian artery. A significant part of stent was hanging into the arch of aorta necessitating removal of stent either surgically or percutaneously. The fully expanded stent was successfully retrieved with the help of Acu-snare and long 8F sheath without any injury or dissection of aorta or iliac artery. After retrieval of stent, the culprit lesion was stented with another self expandable stent in the same sitting with good end result.